

DETAILED TENDER DOCUMENT

**SUPPLY OF LABORATORY EQUIPMENTS FOR
ELECTRICAL ENGINEERING DEPARTMENT**

GOVERNMENT POLYTECHNIC MAYEM
BICHOLIM – GOA

PHONE: 0832-2361235, 2360240

Fax: 0832-2361235

E-mail: - gpbstores@rediffmail.com

Table of contents

E-TENDER NOTICE4-5
TENDER/CATALOGUE DETAILS..... 6-7
TERMS AND CONDITONS.....8- 15
CERTIFICATE..... 16
DOCUMENTS TO BE SUBMITTED ALONGWITH THE TENDERS.....17
DETAILS OF EQUIPMENTS.....18-20

No.GPB/S&P/E-Tender/Electrical/2020/1057
Govt. Polytechnic Mayem,
Bicholim – Goa

Date:17/12/2020

NOTICE INVITING E-TENDER

E-Tenders are invited by The Principal, Govt. Polytechnic Mayem, Bicholim – Goa, on behalf of the Governor of Goa from authorised dealers/manufacturers for the supply of Laboratory Equipments for Electrical Engineering Department

TENDER FEE: Rs.1000/-

E-TENDER PROCESSING FEE: Rs.1500/-

SUPPLY OF LABORATORY EQUIPMENTS FOR ELECTRICAL ENGINEERING DEPARTMENT (TENDER NO. GPB/S&P/E-Tender/Electrical/2020/1057)

| Sr. No. | Description of Laboratory Equipment | Qty. |
|----------------|---|-------------|
| 01. | Cable fault location kit | 01 no. |
| 02. | Meter demonstration | 01 no. |
| 03. | Three phase welding transformer | 01 no. |
| 04. | Electric safety trainer | 01 no. |
| 05. | Home electrical wiring training system | 01 no. |
| 06. | DC Series motor generator set (Coupled) with mounting frames, coupling and decoupling arrangement | 01 no. |
| 07. | 3 Phase Transformer | 06 nos. |
| 08. | 1 phase transformer | 06 nos. |
| 09. | Single Phase Resistive Loading Rheostat | 10 nos. |
| 10. | 1 Phase Auto Transformer | 06 nos |
| 11. | RLC Resonance Trainer | 01 no |
| 12. | Trainer Kit for Three Phase Transformer | 01 no |

- N.B:
1. Specify clearly the details/make of each equipments and also enclose the scanned copies of the product literature online.
 2. The supply, installation and commissioning should be done by the supplier at Govt. Polytechnic Mayem, Bicholim-Goa.
 3. All equipments should carry one year onsite warranty.
 4. The tenderer shall submit authorization letter from manufacturer (Original Equipment manufacturer) for the items mentioned in the tender document.

- All warranty complaints shall be entertained by the vendor.
5. The vendor shall rectify the faults within 15 working days failing which the vendor will arrange temporary (standby) replacement till the faults are rectified. If faults cannot be repaired, permanent replacement should be provided for the same
 6. All warranty complaints shall be entertained by the vendor & registered with the ASC if required.

The detailed tender document can be downloaded from the websites www.goa.gov.in & www.gpb.goa.gov.in Prospective bidders are requested to visit these websites regularly, as any amendment in future will be uploaded to these websites only.

For participation in tender, please visit <https://goenivida.gov.in>

For details on participation in E-Tender Please contact Mr. Sandesh Mandrekar (09881740604) of InfoTech Corporation of Goa Ltd, Altinho, Panajim-Goa.

- i. **Mode of Payment towards Cost of the Tender Document, EMD & Tender Processing Fee:** To be paid online through e-payment mode via NEFT/RTGS/debit card/credit card facility/net banking with pre-printed challans available on e-tendering website and directly credit the amount to ITG account as generated by challans for NEFT/RTGS.
- ii. The Tenderer should use the valid class III Digital Signature Certificate (Signature & Encryption) in order to participate in E-tender.

The tender should be submitted through electronic mode only latest by 3.00 p.m. on **07/01/2021** and will be opened on **11/01/2021** at 11.30 a.m. (Technical Bid) in the presence of tenderers or their authorized representatives, if any. Meanwhile, the tender opening event can be viewed online by logging in to the E-tender site at the time of opening of the bid.

The Principal, Govt. Polytechnic Mayem, Bicholim reserves the right to accept or reject any or all tenders without assigning any reason thereof.

Bicholim – Goa
Date: 17/12/2020

Sd/-
(Subhash P. Borkar)
Principal
Govt. Polytechnic Mayem
Bicholim – Goa

**Government Polytechnic Mayem
Bicholim – Goa**

TENDER / CATALOGUE DETAILS

**GOVERNMENT POLYTECHNIC MAYEM
BICHOLIM – GOA**

TENDER / CATALOGUE DETAILS

| | | |
|----|---|--|
| 1. | Tender Notice | No.GPB/S&P/E-tender/Electrical/2020/1057 |
| 2. | Date of issue of Tender document | 17.12.2020 |
| 3. | Last date of receipt of Tender | 07.01.2021 up to 3.00 P.M. |
| 4. | Date and time of opening tender Technical bid | 11.01.2021 at 11:30 A.M. |
| 5. | Date & Time of Financial bid | To be communicated |
| 6. | Venue of opening tender | Govt. Polytechnic Mayem, Bicholim – Goa |

Sd/-
(Subhash P. Borkar)
Principal
Govt. Polytechnic Mayem
Bicholim – Goa

TERMS AND CONDITIONS

TENDER TERMS AND CONDITIONS
(Subject to legal jurisdiction of Goa)

1) **General Conditions:**

- A. The Principal, Government Polytechnic Mayem, Bicholim reserves the right to reject all or any of the tenders without assigning any reason thereof.
- B. Irrespective of whether mention was made in the Notice inviting tenders or otherwise, only registered dealers under the Sales Tax Act are eligible to submit tenders.
- C. Those who are submitting tenders for the first time, shall invariably enclose documentary evidence in support of the status described at (B) above, failing which such tenders/quotations shall be rejected.
- D. Except where exemption is granted by the Government of Goa, Earnest Money Deposit (E.M.D.) of 2% of the quoted rates of items shall be paid through electronic mode, which will be forfeited if the tenderer derogates, amends or withdraws the offer within validity period.
- E. If the tenderer desires, he/she will be permitted to be present at the time of opening of the tender.
- F. The E.M.D. is refundable in case of unsuccessful tenderer & incase of tenders which are acceptable, the EMD shall be adjusted against the Security Deposit which shall be 10% of the order value.
- G. The tenderer shall not be entitled to any interest/compensation on account of EMD/SD not being refunded in time.
- H. This office will take all the care to return the EMD within the reasonable time limit, which is normally 30 days after opening of the tenders and 30 days for Security Deposit after satisfactory and full execution of the order and on acceptance of the stores by The Principal, Government Polytechnic Mayem, Bicholim. However, if the supplies are covered by warranty period, the SD shall be retained till the warranty period is over. The SD should remain valid for 60 days beyond warranty period.
- I. If the tenderer fails to claim the EMD/SD within the period of 3 months from the time limit specified above, such EMD/SD will be deposited into the Government treasury at the full risk of the supplier and the Principal shall not be responsible for the delay, if any, to reclaim/return the same.
- J. All quotations/tenders will be received for consideration subject to the specified terms and conditions. The tenderers are required to get the necessary clarification regarding any of the terms and conditions before submitting the tender.
- K. Wherever applicable, following details shall be clearly and specifically indicated by the tenderer.
 - i) Disagreement, if any, with terms and conditions enclosed.
 - ii) Non-conformity, if any, of the stores with the specification given in the tender notice/tender paper.

- iii) Taxes, Excise Duty, Custom Duty, etc., if any, to be paid extra.
- iv) Freight, insurance, forwarding and hamali charges, if any, to be paid extra.
- v) The probable suppliers shall be required to demonstrate the equipments/software quoted by them at Govt. Polytechnic Mayem, Bicholim or at the suppliers' place. Expenditure towards the same shall be borne by the vendor.
- vi) Warranty at site shall be one year for equipments, while for computers and related hardware, it shall be three years.
- vii) AMC or extended warranty shall be mentioned separately, wherever required.
- viii) Catalogues/brochures/manuals, test results & certification/calibration certificates to be submitted by the supplier.
- ix) Delivery, cost of installation, loading/unloading and local charges shall be borne by the supplier.
- x) Training at Govt. Polytechnic Mayem, Bicholim shall be provided by the supplier free of cost, wherever required.

L. GST, Excise duty, Custom duty, etc:

Where the cost of the items is specified as exclusive of these statutory levies, unless accepted in writing or otherwise by The Principal, Government Polytechnic Mayem, Bicholim, before placing of the order, the said levies shall be payable only on production of the prescribed documents as below:

- i) Certificate (provisional certificate GST)
- ii) Custom duty certificate issued by the concerned Custom Office.
- iii) Freight & Insurance: Where the cost is exclusive of freight and insurance charges, the tenderer shall invariably provide the documentary evidence for having actually paid such charges, failing which the same shall be disallowed. However, where the rates or amount was not specified in the order, it shall be consistent with current market rates.

M. In case where clarity as at clause (K) above is not adhered to, then The Principal, Government Polytechnic Mayem, Bicholim shall be entitled to and it shall be lawful to do so:

- i) To accept for processing such tenders at the risk of the tenderer and /or
- ii) To reject the tender without assigning any reason thereof.

N. Irrespective of whether specifically stated in the Tender Notice or otherwise, the following tenders will be rejected:

- i) Where the tenderer demands advance payment or payment through bank against invoice or similar form of advance payment.
- ii) Where tender cost, EMD and E-Tender processing fee are not enclosed/paid.
- iii) Where the tenderer has not submitted leaflets, catalogues etc., giving detailed specification of the products.

- O. Unless expressly provided in writing by The Principal, Government Polytechnic Mayem, Bicholim, the order placed shall be subject to the terms and conditions and the specifications given hereinunder. If any of the terms and conditions and /or specifications of the tenderer on whom order is placed is contrary or inconsistent with any of these terms and conditions, the same shall be deemed to be and treated as inapplicable and of no effect.
- P. Any typographical/clerical error in relation to rates/units/quantity, size, etc., and /or contradicting with the tender specifications and /or the terms and conditions shall stand corrected to that effect without any written communication and the tenderer shall not be entitled to any claim whatsoever, based on such typographical/clerical error.
- Q. The tenderer shall sign with date on the certificate attached in token of acceptance of the terms and conditions without any reservations and affix the rubber seal of office stamp, if available. Where rubber stamp is not available, name of the person who has signed and the tenderer which he/she represents shall be clearly written in ink. Quotations received without being duly signed and stamped are liable to be rejected without assigning any reason thereof. However, The Principal, Government Polytechnic Mayem, Bicholim, in the public interest, may relax this condition.
- R. **MODE OF SUBMISSION & LATE TENDERS**
All the tenders are to be submitted in electronic mode only. The tenderwizard will not accept any tender submitted after the deadline for submission of Tenders. All the Tenderers participating in the tender are requested to submit the bids well ahead of the deadline. The Principal, Government Polytechnic Mayem, Bicholim will not accept any responsibility for any problems arising due to delay in the submission of the tender.

2) Validity of Rates:

- A. The rates quoted should be valid at least for a period of 60 days from the date of opening of the tender.
- B. Rates quoted shall be generally exclusive of transport, loading and unloading charges, excise, taxes, etc. However these other charges/taxes shall clearly be shown separately, otherwise it will be deemed that rates quoted are inclusive of all charges/taxes.
- C. The rates shall approximately conform to the current market rates specified by Government/Government approved agencies wherever applicable. If, at any stage, it is discovered that the rates quoted by the tenderer are/were not conforming to the then prevailing market rates, The Principal, Government Polytechnic Mayem, Bicholim reserves the right to reject the stores, if received, in such cases and the tenderer will not have any claim against the Principal, Government Polytechnic Mayem, Bicholim except the actual cost of stores as per normal rates less whatever penalty levied by The Principal, Government Polytechnic Mayem, Bicholim. For this purpose, "Normal rates" means the actual cost of the stores on par with

market rates of similar stores exclusive of taxes, excise, transport, freight, etc., plus reasonable profit margin acceptable as decided by the arbitrator (See Clause no.8).

3) Execution of order:

- A. Acceptance of the order shall be conveyed by the tenderer by submitting a Security Deposit equivalent to 10% of the value of the order, within 15 days from the date of receiving order, in the form of Demand Draft/FDR/Bank Guarantee of any commercial Bank in favour of The Principal, Government Polytechnic Mayem, Bicholim, failing which the order shall stand cancelled, unless informed otherwise in writing by The Principal, Government Polytechnic Mayem, Bicholim. The Demand Draft must be payable at Bicholim.
- B. The Security deposit will be refunded to the supplier after full execution of the order or expiry of the period specified in the order and where the stores are warranted, the S.D. shall be returned after expiry of such warrantee period. If the supplier fails to honour the warrantee as promised, the S.D. will be forfeited.
- C. The entire order shall be executed within three weeks from the date of issue unless specified otherwise, in the order. However, The Principal, Government Polytechnic Mayem, Bicholim reserves the right to grant the extension of the time limit, if in his opinion, public interest does not suffer. Violation of this clause attracts provisions of clauses 5(A) and 5(C).
- D. The stores shall be properly packed and dispatched insured if necessary with any Indian Government approved insurance company or its branch at a cost consistent with Government regulations and the cost of the consignment against loss, damages, breakage, etc., upto destination by Road Transport on freight paid basis only.
- E. It shall be primarily the responsibility of the tenderer to inspect and satisfy that the stores to be supplied conform exactly to the specifications given in the order. In case of lapse on the part of the tenderer in this regard, The Principal, Government Polytechnic Mayem, Bicholim reserves the right to reject the stores.
- F. In case of rejection of the stores supplied by the tenderer on the grounds specified above at D & E or if found defective or not conforming to specification or any other grounds in accordance with the terms and conditions, it shall be the responsibility of the tenderer to make arrangement to collect back the stores. If the tenderer so desires the rejected stores may be dispatched back to the tenderer provided:
 - i) The tenderer gives an undertaking in the prescribed form accepting full responsibility for losses/damages which may be caused during transit due to accidents or any other reasons.
 - ii) The tenderer shall send in advance packing, forwarding and transport charges by way of Demand Draft in favour of The Principal, Government Polytechnic Mayem, Bicholim payable at Bicholim drawn on any Nationalised Bank.

4) **Withdrawal:**

The Successful tenderer cannot refuse to supply the item after the purchase order has been placed & accepted within the validity period. Refusal shall be violation of the Terms & Conditions of this tender, and may lead to blacklisting of the bidder.

5) **The time (delivery period stipulated)** is essence of the contract. In case the Tenderer fails to deliver the stores or any part thereof within the stipulated period of delivery or in case the stores are found not to be in accordance with specifications, The Principal, Government Polytechnic Mayem, Bicholim shall have the right to exercise his discretionary power as under:

- A. Either to recover as liquidated damages a sum not exceeding half percent of the price of stores which the tenderer has failed to deliver as aforesaid per each or part thereof during which the delivery of each may be in arrears but subject to a maximum limit of 5% of the stipulated cost of the stores,
- B. To Purchase from elsewhere at the cost and risk of the tenderer the stores, so undelivered or stores of a similar description without canceling the order in respect of the consignment not yet delivered, or
- C. To cancel the entire/part of this order as deemed fit.

6) **Inspection/Acceptance of Stores:**

- A. Inspection of the stores shall invariably be done at the Polytechnic premises. The stores shall be deposited by the tenderer at his/her risk at the Polytechnic premises. In such cases acknowledgement shall in no case be deemed and will not be deemed as acceptance of the stores.
- B. In case the stores are delivered in packed cases either personally or through transport agencies including Rail/Road agencies, the acknowledgment given by the Storekeeper shall be "On said to contain basis". In such cases, The Principal, Government Polytechnic, Mayem, Bicholim reserves the right to verify actual contents after opening and therefore, the responsibility for damaged/defectives/shortages or the consequences of a similar nature shall be solely the responsibility of the tenderer. Therefore the tenderer may give open delivery by making suitable arrangement.
- C. In cases where the stores are deposited in open container/open condition the acknowledgment of the Storekeeper shall be to the extent of physical quantity and in no case the Storekeeper shall be responsible as to the exact nature/conformity of the stores with the specifications in the order.
- D. The Storekeeper of Government Polytechnic Mayem, Bicholim shall take maximum precaution for the safe custody of stores delivered by the tenderer. However, incase of loss/losses caused to the stores deposited by the tenderer in the Polytechnic premises due to fire, natural calamities or any other reason beyond human control, The Principal, Government Polytechnic Mayem, Bicholim, shall not be responsible to make good any losses, either in full/part or to pay any compensation of whatsoever nature in this regard.

- E. It shall be the responsibility of the Tenderer to arrange delivery of the stores at the Polytechnic premises at Mayem, Bicholim – Goa, during the office hours from 9.30 a.m. to 4.30 p.m.
- F. Advance intimation of dispatch of the stores with information like mode of transport, name and address of transport agency, likely date of arrival of the stores at Polytechnic premises and necessary documents, if any, to release the consignment must be sent by the tenderer to The Principal, Government Polytechnic Mayem, Bicholim.

7) **Warranty/Guarantee:**

- A. Unless specified otherwise in writing the machine/equipment supplied shall carry a warranty/guarantee against manufacturing defects for a period of 12 months from the date of final acceptance of such machine or equipment by The Principal, Government Polytechnic Mayem, Bicholim.
- B. During the warranty/guarantee period the firm responsible for supplying the machine and/or equipment under question shall provide free repairs and servicing, failing which Security Deposit will be forfeited.
- C. During the warranty, if it becomes essential to send/take the equipment to the works/factory of the manufacturer/supplier for replacing and/or servicing, Bank Guarantee in the prescribed form shall be provided by manufacturer/supplier toward the cost of such equipment/machine. The Bank Guarantee so drawn shall be for the period for which equipment remains with the manufacturer/supplier.

8) **Terms of Payment:**

- A. All bills shall be drawn in the name of The Principal, Government Polytechnic Mayem, Bicholim only and shall be submitted in triplicate with the original being affixed with a revenue stamp of Rs.1/- if the bill amounts to Rs.5000/- and above.
- B. Unless accepted otherwise in writing before placing of the order all payments towards the suppliers will be made against valid bills only and if the equipment is found in good condition and conforming to our specifications given in the order.
- C. Normally as per Government rules the payment of bills can be made within a period of 30 days provided the supply reaches Government Polytechnic Mayem, Bicholim in the last week of the month. However, in case of delays, no interest shall be paid by The Principal, Government Polytechnic Mayem, Bicholim on whatsoever any grounds.

9) **Arbitration Clause:**

In the event of there arising any dispute or difference of opinion concerning or touching any clause of this agreement or any of the terms and conditions in any of the documents which are made integral part of this agreement, the same shall be referred to the sole arbitration of a person appointed by the Governor of Goa and it shall be lawful for the Governor of Goa to appoint any officer who has in the discharge of his

duties as such officer either expressed any opinion or issued any direction in the matter shall be final & binding on the parties.

Sd/-
Principal
Govt. Polytechnic Mayem,
Bicholim – Goa

CERTIFICATE

Certified that I/we have gone through the terms and conditions of supply/works and also the tender conditions. I/We hereby agree to abide by the same. In token of accepting as above I/We affix the signature as below:

Signature

Date: _____

Seal of the Tenderer

Name of the Firm/Organisation: _____

Contact Person: _____

Tel No: _____

Mobile No.: _____

E-mail ID: _____

NOTE: This certificate duly filled and signed shall be uploaded alongwith the tender offer without which the tender will not be considered.

DOCUMENTS TO BE SUBMITTED ALONGWITH THE TENDERS

- a) Certificate of Acceptance of terms and conditions duly filled and signed.
- b) Income Tax returns of current year.
- c) CGST/SGST Clearance certificate.
- d) The Tenderer should have documents supporting that he is manufacturer/ authorized dealer.
- e) Name and details of local agency who will attend to the complaints, in case required.
- f) Scanned copies of the product literature

DETAILS OF EQUIPMENTS

ANNEXURE – I

| SR.NO | Description of Item | Qty. |
|-------|--|--------|
| 01. | <p>Cable fault location kit: with distance, resistance measurement features and metering arrangement: including following options for fault location using megger; varley loop method, murray loop method; pulse reflection method with provision of suitable length 3.5 core cable</p> | 01 no. |
| 02. | <p>Meter demonstration: for study of operation of moving coil, moving iron, dynamometer type instrument setup includes: experiment panel that contains one each following meters : wattmeter, voltmeter(mc, mi type), ammeter(mc, mi type), ac load connector, dc load connector, connecting patch cords, learning material CD</p> | 01 no. |
| 03. | <p>Three phase welding transformer: Output current: 60-400amps, output voltage: 65 volts, cooling air forced, input current: 50amps, supply voltage: 380-415volts ac with suitable current controller</p> | 01 no. |
| 04. | <p>Electric safety trainer : for performing different electrical activities to avoid electrical hazards and for importance of earthing; to study of role of fuse in any electrical / electronic circuits; for working of mcb, mccb set up should include hair dryer as a load, fuses; experiment panel that contains red and green Leds, mcb, mccb, ammeter, connecting patch cord, learning material CD</p> | 01 no. |
| 05. | <p>Home electrical wiring training system : for study of connection / measurement of energy meter and consumer unit, voltage and currents, mcb wiring, tube light wiring, two way switch wiring, short circuit fault, switchboard ceiling fan, series and parallel operation on load setup includes: ceiling fan , experimental panel that contains graphical lcd, load connectors, choke coil for tubelight, mcb,connecting patch cords, learning material CD.</p> | 01 no. |
| 06. | <p>DC Series motor generator set (Coupled) with mounting frames, coupling and decoupling arrangement: dc series motor: fields: series; rating 5hp; 230v dc for armature and field , 1500 rpm speed; current rating 20 amps with braking arrangement for motor alone DC series generator: fields: series; rating 3kw; 230v dc for armature and field , 1500 rpm speed; current rating 15 amps</p> | 01 no. |

| | | |
|----|--|---------|
| | <p>panel: separate current and voltage metering for motor & generator ; drum controller starter with heavy duty banana terminals mounted on panel; facility for input and output connection through heavy duty banana terminals</p> <p>load: 2.5 kw, 230 volts, 10 amps load variable in 10 steps through heavy duty switches; current and voltage metering features and protection equipment</p> | |
| 07 | <p>3 phase transformer: 3kva;50hz;415 volts input AC;230 volts output AC; tapping at 75,115 volts AC on secondary; heavy duty banana terminals for input and output connections; star /delta connection</p> | 06 nos. |
| 08 | <p>1 phase transformer : 1Kva;50hz 230 volts input ac; 115 volts output ac; tapping at 50 v, 100v,115 volts ac on secondary; heavy duty banana terminals for input and output connections</p> | 06 nos. |
| 09 | <p>Single phase resistive loading rheostat: Rheostat 2.5 kw, 10amps, 230 volts, variable in 10 steps through heavy duty switch with current and voltage metering features and protection equipment.</p> | 10 nos. |
| 10 | <p>1 phase auto transformer : 10amps; 230 volts input AC; (0-270) volts output AC; voltage variable knob brought out and heavy duty banana terminals for input and output connections; voltage metering facility; close type desk top model with on indicator.</p> | 06 nos. |
| 11 | <p>RLC Resonance Trainer : for study of RLC series and parallel resonance circuit and to determine its resonance frequency unit has on board function generator, frequency counter, LCD for frequency and voltage counter reading</p> | 01 no. |
| 12 | <p>Trainer kit for Three phase transformer : includes: experiment card with three phase transformer, with 12 windings and tapping for study of single phase and three phase transformer and transformer circuits, three phase load, usable for star and delta connection; CD rom with labsoft and course. Course content: Principal of transformer study of load characteristics of single phase transformer, for one quadrant and four quadrant operation measurement of currents and voltage under load/no load condition study of transformation ration equivalent circuit diagram study of three phase transformer study of various three phase transformer circuit and their effects on load/ no load operation study of various circuit with unbalance load determination of short circuit</p> | 01 no. |

| | | |
|--|---------------------------|--|
| | voltage duration:3h appx. | |
|--|---------------------------|--|